

hs-CRP Calibrator Set

use for Measure hs-CRP reagent

Intended Use

This product is for purpose of calibrating results in the quantitative determination of C-reactive protein in Serum or Plasma automated immunoassay. For in vitro diagnostic use only.

Precautions

HBs Ag : Negative (—)
HCV Ab : Negative (—)
HIV Ab : Negative (—)

The product must however be handled as potentially biohazardous material.

This product includes 0.09w/v% sodium azide as preservative.

In case the product gets into eye or mouth, or comes into contact with skin, wash with plenty of running water and consult a doctor if needed.

Storage : Dark and Cold (2~8 °C)

Lot No. : 123ACI

Exp. Date : 2025. 03.

【Assigned value】

Level	Color (cap)	mg/dL	mg/L
1	Colorless	0.00	0.0
2	Blue	0.16	1.6
3	Green	1.50	15.0
4	Yellow	4.55	45.5
5	Orange	18.5	185
6	Pink	38.0	380

Manufactured by :

UMA Co., Ltd.

2-19-6 Yokosuka Matsudo City Chiba-Ken,

270-0031 Japan

phone: +81-47-710-4871

試験者	品質管理課長	製造管理者
		