

## hs-CRP Control Set

use for Measure CRP, Measure CRP-a and Measure hs-CRP reagent

### Intended Use

This product is for purpose of calibrating results in the quantitative determination of C-reactive protein in Serum or Plasma automated immunoassay. For in vitro diagnostic use only.

### Precautions

|          |          |     |
|----------|----------|-----|
| HBs Ag : | Negative | (-) |
| HCV Ab : | Negative | (-) |
| HIV Ab : | Negative | (-) |

The product must however be handled as potentially biohazardous material.

This product includes 0.09w/v% sodium azide as preservative.

In case the product gets into eye or mouth, or comes into contact with skin, wash with running water and consult a doctor if needed.

Storage : Dark and Cold (2~8 °C)

Lot No. : 116ACB

Exp. Date : 2024.08.

### 【Control value】

| Level | Color (cap) | Measured value |       | Reference range |
|-------|-------------|----------------|-------|-----------------|
| Low   | Purple      | 1.15           | mg/dL | 0.85 - 1.45     |
|       |             | 11.50          | mg/L  |                 |
| High  | Brown       | 2.44           | mg/dL | 2.14 - 2.74     |
|       |             | 24.40          | mg/L  |                 |

manufactured by :

UMA Co., Ltd.

2-19-6 Yokosuka Matsudo City Chiba-Ken,

270-0031, Japan

phone: +81-47-710-4871

| 試験者   | 品質管理課長  | 製造管理者   |
|---|---|---|
|  |  |  |