

# hs-CRP Calibrator Set

use for Measure hs-CRP reagent

## Intended Use

This product is for purpose of calibrating results in the quantitative determination of C-reactive protein in Serum or Plasma automated immunoassay. For in vitro diagnostic use only.

## Precautions

HBs Ag : Negative (-)  
HCV Ab : Negative (-)  
HIV Ab : Negative (-)

The product must however be handled as potentially biohazardous material.

This product includes 0.09w/v% sodium azide as preservative.

In case the product gets into eye or mouth, or comes into contact with skin, wash with plenty of running water and consult a doctor if needed.

Storage : Dark and Cold (2~8 °C)

Lot No. : 124ADD

Exp. Date : 2025. 10.

### 【Assigned value】

Level	Color (cap)	mg/dL	mg/L
1	Colorless	0.00	0.0
2	Blue	0.16	1.6
3	Green	1.60	16.0
4	Yellow	4.70	47.0
5	Orange	18.5	185
6	Pink	38.5	385

Manufactured by :

UMA Co., Ltd.

2-19-6 Yokosuka Matsudo City Chiba-Ken,

270-0031 Japan

phone:+81-47-710-4871

試験者	品質管理課長	製造管理者
		