

Lipids Calibrator Certificate of Analysis

Issued Date : 2024/12/2
 Certificate No. : 330
 Lot No. : 121ADL
 Expiration date : 2026/06

Origin : Freeze Dried Pooled Human Serum
 each specimen has passed HBV , HCV and HIV tests.

Table 1, Certified Concentrations for Lipid Calibrator

Analyte		Standard	Result			
HDL-C	HDL	≥ 50 mg/dL	55.0	mg/dL	1.42	mmol/L
LDL-C	LDL		113	mg/dL	2.92	mmol/L
TC			208	mg/dL	5.38	mmol/L
TG			131	mg/dL	1.48	mmol/L
T-TG			131	mg/dL	1.48	mmol/L
HBV Ag.Ab		Negative (—)	Negative (—)			
HCV Ag.Ab		Negative (—)	Negative (—)			
HIV Ag.Ab		Negative (—)	Negative (—)			
Tp		Negative (—)	Negative (—)			
ATLV Ab.		Negative (—)	Negative (—)			

Preparation : Reconstitute 2 mL of distilled water and mix gently,
 then store room temperature and sometimes mix
 during 2 hours before use.

Measurement by : Takekazu Yoshitani
 Date : 2024/12/2