

試験成績書
Certificate of Analysis

株式会社ユーエムエー
UMA Co.,Ltd

試験書番号 Certificate No.	1057
コード Code No.	21D601~606
製品名 Product	D-dimer Calibrator (STD)
試験実施年月日 Examination date	Year MM DD 2024/12/2
総合判定 Result	合格 Accepted

Examination Reagent Lot 106ADL

Manufactured Date 2024/12/2
Year/MM/DD

Expiry Date 2026/06
有効期限 Year/MM

Analytes	Result
HBsAg	negateve(-)
HIVab	negateve(-)
HCVab	negateve(-)
Calib.-1 (*WHITE)	0.0 μ g/mL
Calib.-2 (*BLUE)	0.52 μ g/mL
Calib.-3 (*GREEN)	1.3 μ g/mL
Calib.-4 (*YELLOW)	5.9 μ g/mL
Calib.-5 (*ORANGE)	14.2 μ g/mL
Calib.-6 (*PINK)	36.5 μ g/mL

* Color of cap

* From this lot, the dissolution amount will be changed to 1.0 mL / vial.
* 本ロットより、溶解量が1.0mL/vialに変更となります。

試験者	品質管理課長	製造管理者
		