

Bilirubin Calibrator Certificate of Analysis

Issued Date : 2025/12/1
Certificate No. : 15
Lot No. : 009AEL
Manufacture date : 2025/12/1
Expiration date : 2027/06

Origin : Freeze Dried Pooled Human Serum
each specimen has passed HBV , HCV and HIV tests.

Table 1, Certified Concentrations for bilirubin Calibrator

Analyte	Standard	Result
T-BIL		12.0 mg/dL
D-BIL		9.4 mg/dL
		mg/dL
		mg/dL
		mg/dL
		mg/dL
		mg/dL
		mg/dL
HBV Ag.Ab	Negative (—)	Negative (—)
HCV Ag.Ab	Negative (—)	Negative (—)
HIV Ag.Ab	Negative (—)	Negative (—)
Tp	Negative (—)	Negative (—)
ATLV Ab.	Negative (—)	Negative (—)

Preparation : Reconstitute 2 mL of distilled water and
then stand for 20 minutes, swirling occasionally.
Store in the dark after reconstitution.

Measurement by : Takeshi Yoshida
Date : 2025/12/1