

Lipid Calibrator Certificate of Analysis

Issued Date : 2025/6/2
 Certificate No. : 336
 Lot No. : 123AEF
 Expiration date : 2026/12

Origin : Freeze Dried Pooled Human Serum
 each specimen has passed HBV , HCV and HIV tests.

Table 1, Certified Concentrations for Lipid Calibrator

Analyte		Standard	Result	
HDL-C	HDL	≥ 50 mg/dL	64.0 mg/dL	1.66 mmol/L
LDL-C	LDL		116 mg/dL	3.00 mmol/L
TC			210 mg/dL	5.43 mmol/L
TG			124 mg/dL	1.40 mmol/L
T-TG			125 mg/dL	1.41 mmol/L
HBV Ag.Ab		Negative (—)	Negative (—)	
HCV Ag.Ab		Negative (—)	Negative (—)	
HIV Ag.Ab		Negative (—)	Negative (—)	
Tp		Negative (—)	Negative (—)	
ATLV Ab.		Negative (—)	Negative (—)	

Preparation : Reconstitute 2 mL of distilled water and mix gently,
 then store room temperature and sometimes mix
 during 2 hours before use.

Measurement by : Takeshi Yoshida
 Date : 2025/6/2