

Lipid Control Certificate of Analysis

Issued Date : 2025/12/1
 Certificate No. : 344
 Lot No. : 124AEL
 Expiration date : 2027/06

Origin : Freeze Dried Pooled Human Serum
 each specimen has passed HBV , HCV and HIV tests.

Table 1, Certified Concentrations for Lipid Calibrator

Analyte		Standard	Result (Reference Range)		
HDL-C	HDL		65.0	59.0-71.0	mg/dL
			1.68	1.53-1.84	mmol/L
LDL-C	LDL		118.0	105-131	mg/dL
			3.05	2.72-3.39	mmol/L
TC			207.0	185-229	mg/dL
			5.35	4.78-5.92	mmol/L
TG (T-TG)			123.0	111-185	mg/dL
			1.39	1.25-1.52	mmol/L
			125.0	(113-137)	mg/dL
			1.41	(1.28-1.55)	mmol/L
HBV Ag.Ab		Negative (—)	Negative (—)		
HCV Ag.Ab		Negative (—)	Negative (—)		
HIV Ag.Ab		Negative (—)	Negative (—)		
Tp		Negative (—)	Negative (—)		
ATLV Ab.		Negative (—)	Negative (—)		

Preparation : Reconstitute 2 mL of distilled water and mix gently,
 then store room temperature and sometimes mix
 during 2 hours before use.

Measurement by : Takeshi Yoshida
 Date : 2025/12/1

UMA Co., Ltd. : 2-19-6 Yokosuka Matsudo City Chiba-Ken, 270-0031, Japan