

PG I /PG II Control

PG I /PG II Control Set

Intended Use

Control for PG I /PG II

Precautions

HBs Ag :	Negative	(-)
HCV Ab :	Negative	(-)
HIV Ab :	Negative	(-)

The product must however be handled as potentially biohazardous material.

This product includes 0.09w/v% sodium azide as preservative.

After the measurement, it must be diluted with plenty of water and then drained away.

Storage : Dark and Cold (2~8 °C)

Lot No. : 138AEC

Exp. Date : 2026.09.

【Control value】

Level	Color (cap)	Measured value (ng/mL)	Reference range
PG I	LOW Purple	49.1	39.1 - 59.1
	High Brown	150.7	142.7 - 158.7
PG II	LOW Purple	19.0	13.0 - 25.0
	High Brown	68.5	61.5 - 75.5

manufactured by :

UMA Co., Ltd.

2-19-6 Yokosuka Matsudo City Chiba-Ken,

270-0031, Japan

phone: +81-47-710-4871

試験者	品質管理課長	製造管理者
		