hs-CRP Control Set

use for Measure CRP, Measure CRP-a and Measure hs-CRP reagent

Intended Use

This product is for purpose of calibrating results in the quantitave determination of C-reactive protein in Serum or Plusma automated immunoassay. For in vitro diagnostic use only.

Precautions

 HBs Ag :
 Negative
 (-)

 HCV Ab :
 Negative
 (-)

 HIV Ab :
 Negative
 (-)

The product must however be handled as potentially biohazardous material.

This product includes 0.09w/v% sodium azide as preservative. In case the product gets into eye or mouth, or comes into contact with skin, wa of running water and consult a doctor if needed.

Storage : Dark and Cold (2∼8 °C)

Lot No.: 113AAL Exp. Date: 2023.06.

[Control value]

(2021/12/16 reassign)

Level	Color (cap)	Measured value		Reference range
Low	Purple	1. 21	mg/dL	0. 91 - 1. 51
		12. 10	mg/L	9. 10 - 15. 10
High	Brown	2. 54	mg/dL	2. 24 - 2. 84
		25. 40	mg/L	22. 40 - 28. 40

manufactured by :

UMA Co., Ltd.

2-19-6 Yokosuka Matsudo City Chiba-Ken, 270-0031, Japan

phone: +81-47-710-4871

試験者	品質管理課長	製造管理者
(E)	(E)	(T)
海	(A)	(題)