

hs-CRP Control Set

use for Measure CRP, Measure CRP-a and Measure hs-CRP reagent

Intended Use

This product is for purpose of calibrating results in the quantitative determination of C-reactive protein in Serum or Plasma automated immunoassay. For in vitro diagnostic use only.

Precautions

HBs Ag :	Negative	(-)
HCV Ab :	Negative	(-)
HIV Ab :	Negative	(-)

The product must however be handled as potentially biohazardous material.

This product includes 0.09w/v% sodium azide as preservative.

In case the product gets into eye or mouth, or comes into contact with skin, wash with running water and consult a doctor if needed.

Storage : Dark and Cold (2~8 °C)

Lot No. : 118ACF

Manufacture Date : 2023.06.01.

Exp. Date : 2024.12.

【Control value】

Level	Color (cap)	Measured value		Reference range
Low	Purple	1.14	mg/dL	0.84 - 1.44
		11.40	mg/L	8.4 - 14.4
High	Brown	2.58	mg/dL	2.28 - 2.88
		25.80	mg/L	22.8 - 28.8

manufactured by :

UMA Co., Ltd.

2-19-6 Yokosuka Matsudo City Chiba-Ken,

270-0031, Japan

phone: +81-47-710-4871

試験者	品質管理課長	製造管理者
		