

# Lipids Calibrator Certificate of Analysis

Issued Date : 2023/5/1  
 Certificate No. : 308  
 Lot No. : 120ACE  
 Manufacture date : 2023/5/1  
 Expiration date : 2024/11


Origin : Freeze Dried Pooled Human Serum  
 each specimen has passed HBV , HCV and HIV tests.

Table 1, Certified Concentrations for Lipid Calibrator

Analyte		Standard	Result	
HDL-C	HDL	$\geq 50$ mg/dL	59.0 mg/dL	1.53 mmol/L
LDL-C	LDL		136 mg/dL	3.52 mmol/L
TC			210 mg/dL	5.43 mmol/L
TG			118 mg/dL	1.33 mmol/L
T-TG			120 mg/dL	1.35 mmol/L
HBV Ag.Ab		Negative (-)	Negative (-)	
HCV Ag.Ab		Negative (-)	Negative (-)	
HIV Ag.Ab		Negative (-)	Negative (-)	
TP		Negative (-)	Negative (-)	
ATLV Ab.		Negative (-)	Negative (-)	

**Preparation** : Reconstitute 2 mL of distilled water and mix gently, then store room temperature and sometimes mix during 2 hours before use.

**Measurement by** :  
**Date** :

  
 -----  
 2023/5/1  
 -----