

hs-CRP Control Set

use for Measure CRP, Measure CRP-a and Measure hs-CRP reagent

Intended Use

This product is for purpose of calibrating results in the quantitative determination of C-reactive protein in Serum or Plasma automated immunoassay. For in vitro diagnostic use only.

Precautions

HBs Ag :	Negative	(-)
HCV Ab :	Negative	(-)
HIV Ab :	Negative	(-)

The product must however be handled as potentially biohazardous material.

This product includes 0.09w/v% sodium azide as preservative.

In case the product gets into eye or mouth, or comes into contact with skin, wash with running water and consult a doctor if needed.

Storage : Dark and Cold (2~8 °C)

Lot No. : 121ADJ

Exp. Date : 2026.04.

【Control value】

Level	Color (cap)	Measured value		Reference range
Low	Purple	1.10	mg/dL	0.8 - 1.4
		11.00	mg/L	8 - 14
High	Brown	2.69	mg/dL	2.39 - 2.99
		26.90	mg/L	23.9 - 29.9

manufactured by :

UMA Co., Ltd.

2-19-6 Yokosuka Matsudo City Chiba-Ken,

270-0031, Japan

phone: +81-47-710-4871

試験者	品質管理課長	製造管理者
		