

Lipids Calibrator Certificate of Analysis

Issued Date : 2024/10/1
 Certificate No. : 328
 Lot No. : 121ADJ
 Expiration date : 2026/04

Origin : Freeze Dried Pooled Human Serum
 each specimen has passed HBV , HCV and HIV tests.

Table 1, Certified Concentrations for Lipid Calibrator

Analyte		Standard		Result			
HDL-C	HDL	≥ 50	mg/dL	55.0	mg/dL	1.42	mmol/L
LDL-C	LDL			113	mg/dL	2.92	mmol/L
TC				208	mg/dL	5.38	mmol/L
TG				131	mg/dL	1.48	mmol/L
T-TG				131	mg/dL	1.48	mmol/L
HBV Ag.Ab		Negative (-)		Negative (-)			
HCV Ag.Ab		Negative (-)		Negative (-)			
HIV Ag.Ab		Negative (-)		Negative (-)			
TP		Negative (-)		Negative (-)			
ATLV Ab.		Negative (-)		Negative (-)			

Preparation : Reconstitute 2 mL of distilled water and mix gently,
 then store room temperature and sometimes mix
 during 2 hours before use.

Measurement by :
 Date :

Takekazu Yoshida

 2024/10/1
