Bilirubin Calibrator Certificate of Analysis

Issued Date	:	2025/5/1
Certificate No.	:	13
Lot No.	:	009AEE
Expiration date	:	2026/11

Origin : Freeze Dried Pooled Human Serum
each specimen has passed HBV , HCV and HIV tests.

Table 1, Certified Concentrations for bilirubin Calibrator

Analyte	Standard	Result
T-BIL		12.0 mg/dL
D-BIL		9.4 mg/dL
		mg/dL
		mg/dL
		mg/dL
		mg/dL
		mg/dL
		mg/dL
HBV Ag.Ab	Negative (−)	Negative (−)
HCV Ag.Ab	Negative (−)	Negative (−)
HIV Ag.Ab	Negative (−)	Negative (−)
Tp	Negative (−)	Negative (−)
ATLV Ab.	Negative (−)	Negative (−)

Preparation : **Reconstitute 2 mL** of distilled water and then stand for 20 minutes, swirling occasionally. Store in the dark after reconstitution.

Measurement by : Takeuchi

Date : 2025/5/1

UMA Co., Ltd. :2-19-6 Yokosuka Matsudo City Chiba-Ken,270-0031, Japan