

# Bilirubin Calibrator Certificate of Analysis

Issued Date : 2025/12/1  
 Certificate No. : 15  
 Lot No. : 009AEL  
 Manufacture date : 2025/12/1  
 Expiration date : 2027/06

Origin : Freeze Dried Pooled Human Serum  
 each specimen has passed HBV , HCV and HIV tests.

Table 1, Certified Concentrations for bilirubin Calibrator

Analyte	Standard	Result
T-BIL		12.0 mg/dL
D-BIL		9.4 mg/dL
		mg/dL
		mg/dL
		mg/dL
		mg/dL
		mg/dL
		mg/dL
HBV Ag.Ab	Negative (—)	Negative (—)
HCV Ag.Ab	Negative (—)	Negative (—)
HIV Ag.Ab	Negative (—)	Negative (—)
Tp	Negative (—)	Negative (—)
ATLV Ab.	Negative (—)	Negative (—)

**Preparation** : Reconstitute 2 mL of distilled water and then stand for 20 minutes, swirling occasionally. Store in the dark after reconstitution.

Measurement by : Takeshi Yoshida  
 Date : 2025/12/1