

Lipid Control Certificate of Analysis

Issued Date : 2025/12/1
 Certificate No. : 344
 Lot No. : 124AEL
 Expiration date : 2027/06

Origin : Freeze Dried Pooled Human Serum
 each specimen has passed HBV , HCV and HIV tests.

Table 1, Certified Concentrations for Lipid Calibrator

Analyte	Standard	Result (Reference Range)
HDL-C HDL		65.0 59.0-71.0 mg/dL
		1.68 1.53-1.84 mmol/L
LDL-C LDL		118.0 105-131 mg/dL
		3.05 2.72-3.39 mmol/L
TC		207.0 185-229 mg/dL
		5.35 4.78-5.92 mmol/L
TG (T-TG)		123.0 111-185 mg/dL
		1.39 1.25-1.52 mmol/L
		125.0 (113-137) mg/dL
		1.41 (1.28-1.55) mmol/L
HBV Ag.Ab	Negative (-)	Negative (-)
HCV Ag.Ab	Negative (-)	Negative (-)
HIV Ag.Ab	Negative (-)	Negative (-)
Tp	Negative (-)	Negative (-)
ATLV Ab.	Negative (-)	Negative (-)

Preparation : Reconstitute 2 mL of distilled water and mix gently,
 then store room temperature and sometimes mix
 during 2 hours before use.

Measurement by : Takeshi Yoshida
 Date : 2025/12/1

UMA Co., Ltd. : 2-19-6 Yokosuka Matsudo City Chiba-Ken, 270-0031, Japan